

Continuous Radon Monitor

Model Number: 1028
Calibration Date: 09/28/2012
Monitor Time: 12/27/2012 15:15

Serial Number: 49745178
CF: 2.84

Inspector Information

Stephen R. Burns, Inc.
P.O. Box 943
Iowa City
IA-52244
Phone Number: 319-400-4844
License Number: S00227

Customer Information


Wood Elementary
Room 163
1930 Lakeside DR
Iowa City
IA-52240

Site & Condition

Wind:	NA	Atmospheric Condition:	NA
Year Built:	NA	Structure Type:	NA
SqFt:	NA	Monitor Location:	NA

Test Summary

Start Time:	12/24/2012 09:15	Overall Avg:	0.6 pCi/l
End Time:	12/26/2012 09:15	EPA Avg:	0.6 pCi/l
Measurement Interval(hr):	1.0		
Elapse Time:	2 Days 0 hrs		



***** 12/24/2012 *****

Time	Counts pCi/l	Flags
10:15	0.4	
11:15	0.4	
12:15	1.1	
13:15	1.8	
14:15	1.1	
15:15	0.7	
16:15	0.0	
17:15	1.1	
18:15	1.1	
19:15	1.4	
20:15	0.7	
21:15	0.7	
22:15	1.4	
23:15	0.4	

***** 12/25/2012 *****

Time	Counts pCi/l	Flags
00:15	0.4	
01:15	0.0	
02:15	0.4	
03:15	0.0	
04:15	1.1	
05:15	0.0	
06:15	0.4	
07:15	0.7	
08:15	1.1	
09:15	0.4	
10:15	0.4	
11:15	0.4	
12:15	1.1	
13:15	0.7	
14:15	0.0	
15:15	0.0	
16:15	0.4	
17:15	0.7	
18:15	1.1	
19:15	0.4	
20:15	0.0	
21:15	0.7	
22:15	0.4	
23:15	0.0	

***** 12/26/2012 *****

Time	Counts pCi/l	Flags
00:15	0.4	

***** 12/26/2012 *****

Time	Counts pCi/l	Flags
01:15	1.1	
02:15	0.7	
03:15	1.1	
04:15	0.4	
05:15	1.1	
06:15	0.4	
07:15	0.4	
08:15	0.7	
09:15	2.1	

Error Flags:

M Motion:

Inspector Signature _____

PC Software Version: 2.1.1.0

Embedded Software Version: 108