

Continuous Radon Monitor

Model Number: 1028
Calibration Date: 09/28/2012
Monitor Time: 12/29/2012 17:17

Serial Number: 59237107
CF: 2.88

Inspector Information

Stephen R. Burns Inc.
P.O. BOX 943
Iowa City
IA-52244
Phone Number: 319-400-4844
License Number: S00227

Customer Information

Wendell Johnson Kid Talk
Room 229
250 Hawkins DR
Iowa City
IA-52242

Site & Condition


Wind: NA
Year Built: NA
SqFt: NA

Atmospheric Condition: NA
Structure Type: NA
Monitor Location: NA

Test Summary

Start Time: 12/26/2012 13:25
End Time: 12/28/2012 13:25
Measurement Interval(hr): 1.0
Elapse Time: 2 Days 0 hrs

Overall Avg: 0.4 pCi/l
EPA Avg: 0.4 pCi/l



***** 12/26/2012 *****

Time	Counts pCi/l	Flags
14:25	0.3	
15:25	0.0	
16:25	0.3	
17:25	1.4	
18:25	0.3	
19:25	0.0	
20:25	0.3	
21:25	0.0	
22:25	0.3	
23:25	0.0	

***** 12/27/2012 *****

Time	Counts pCi/l	Flags
00:25	0.3	
01:25	0.7	
02:25	1.0	
03:25	0.0	
04:25	0.0	
05:25	0.0	
06:25	0.7	
07:25	0.0	
08:25	0.3	
09:25	0.3	
10:25	0.3	
11:25	0.3	
12:25	0.0	
13:25	1.0	
14:25	0.3	
15:25	0.0	
16:25	0.7	
17:25	0.0	
18:25	0.3	
19:25	1.0	
20:25	0.7	
21:25	0.7	
22:25	0.3	
23:25	0.0	

***** 12/28/2012 *****

Time	Counts pCi/l	Flags
00:25	0.3	
01:25	0.7	
02:25	0.0	
03:25	0.7	

***** 12/28/2012 *****

Time	Counts pCi/l	Flags
04:25	0.0	
05:25	0.3	
06:25	1.4	
07:25	0.7	
08:25	0.7	
09:25	0.3	
10:25	0.0	
11:25	0.3	
12:25	0.0	
13:25	0.0	

Error Flags:

M Motion:

Inspector Signature _____

PC Software Version: 2.1.1.0

Embedded Software Version: 108