

Continuous Radon Monitor

Model Number: 1028
Calibration Date: 02/24/2012
Monitor Time: 12/29/2012 14:14

Serial Number: 51446083
CF: 2.95

Inspector Information

Stephen R. Burns, Inc.
P.O. Box 943
Iowa City
IA-52244
Phone Number: 319-400-4844
License Number: S00227

Customer Information


Wendell Johnson Kid Talk
Room 219
250 Hawkins DR
Iowa City
IA-52242

Site & Condition

Wind: NA
Year Built: NA
SqFt: NA
Atmospheric Condition: NA
Structure Type: NA
Monitor Location: NA

Test Summary

Start Time: 12/26/2012 11:00
End Time: 12/28/2012 11:00
Measurement Interval(hr): 1.0
Elapse Time: 2 Days 0 hrs
Overall Avg: 0.3 pCi/l
EPA Avg: 0.3 pCi/l



***** 12/26/2012 *****

Time	Counts pCi/l	Flags
12:00	0.3	
13:00	0.7	
14:00	0.7	
15:00	0.0	
16:00	1.0	
17:00	0.3	
18:00	0.3	
19:00	0.3	
20:00	0.0	
21:00	0.0	
22:00	0.0	
23:00	0.0	

***** 12/27/2012 *****

Time	Counts pCi/l	Flags
00:00	1.0	
01:00	0.3	
02:00	0.0	
03:00	1.0	
04:00	0.0	
05:00	0.0	
06:00	0.0	
07:00	0.3	
08:00	0.7	
09:00	0.3	
10:00	0.3	
11:00	0.7	
12:00	0.7	
13:00	1.0	
14:00	0.3	
15:00	0.0	
16:00	0.3	
17:00	0.3	
18:00	0.0	
19:00	0.7	
20:00	0.0	
21:00	0.3	
22:00	1.0	
23:00	0.0	

***** 12/28/2012 *****

Time	Counts pCi/l	Flags
00:00	0.7	
01:00	0.3	

***** 12/28/2012 *****

Time	Counts pCi/l	Flags
02:00	0.0	
03:00	0.0	
04:00	0.7	
05:00	0.3	
06:00	0.0	
07:00	0.3	
08:00	0.0	
09:00	0.0	
10:00	0.3	
11:00	0.0	

Error Flags:

M Motion:

Inspector Signature _____

PC Software Version: 2.1.1.0

Embedded Software Version: 108