

Continuous Radon Monitor

Model Number: 1028
Calibration Date: 09/28/2012
Monitor Time: 12/23/2012 22:22

Serial Number: 49745178
CF: 2.84

Inspector Information

Stephen R. Burns, Inc.
P.O. Box 943
Iowa City
IA-52244
Phone Number: 319-400-4844
License Number: S00227

Customer Information


Shimek Elementary
Room 10
1400 Grissel PL
Iowa City
IA-52245

Site & Condition

Wind:	NA	Atmospheric Condition:	NA
Year Built:	NA	Structure Type:	NA
SqFt:	NA	Monitor Location:	NA

Test Summary

Start Time:	12/21/2012 16:29	Overall Avg:	1.2 pCi/l
End Time:	12/23/2012 16:29	EPA Avg:	1.3 pCi/l
Measurement Interval(hr):	1.0		
Elapse Time:	2 Days 0 hrs		



***** 12/21/2012 *****

Time	Counts pCi/l	Flags
17:29	0.0	
18:29	0.7	
19:29	1.4	
20:29	0.7	
21:29	1.4	
22:29	0.4	
23:29	1.1	

***** 12/22/2012 *****

Time	Counts pCi/l	Flags
00:29	2.5	
01:29	1.4	
02:29	1.4	
03:29	1.1	
04:29	1.4	
05:29	0.7	
06:29	1.4	
07:29	0.7	
08:29	2.5	
09:29	0.7	
10:29	2.1	
11:29	1.4	
12:29	2.5	
13:29	2.5	
14:29	2.5	
15:29	0.7	
16:29	1.1	
17:29	0.7	
18:29	1.1	
19:29	1.4	
20:29	2.8	
21:29	1.8	
22:29	0.7	
23:29	1.8	

***** 12/23/2012 *****

Time	Counts pCi/l	Flags
00:29	0.4	
01:29	1.4	
02:29	2.5	
03:29	0.7	
04:29	1.4	
05:29	1.4	
06:29	0.7	

***** 12/23/2012 *****

Time	Counts pCi/l	Flags
07:29	1.4	
08:29	1.4	
09:29	1.8	
10:29	0.0	
11:29	0.4	
12:29	0.0	
13:29	0.7	
14:29	1.1	
15:29	0.7	
16:29	0.4	

Error Flags:

M Motion:

Inspector Signature _____

PC Software Version: 2.1.1.0

Embedded Software Version: 108