

Continuous Radon Monitor

Model Number: 1028
Calibration Date: 09/28/2012
Monitor Time: 1/14/2013 8:8

Serial Number: 57077193
CF: 2.97

Inspector Information

Stephen R. Burns, Inc
P.O. Box 943
Iowa City
IA-52244
Phone Number: 319-400-4844
License Number: S00227

Customer Information


Shimek Elementary
Room 9
1400 Grissel PL
Iowa City
IA-52245

Site & Condition

Wind: NA
Year Built: NA
SqFt: NA
Atmospheric Condition: NA
Structure Type: NA
Monitor Location: NA

Test Summary

Start Time: 01/11/2013 16:46
End Time: 01/13/2013 16:46
Measurement Interval(hr): 1.0
Elapse Time: 2 Days 0 hrs
Overall Avg: 3.9 pCi/l
EPA Avg: 4.1 pCi/l



***** 01/11/2013 *****

Time	Counts pCi/l	Flags
17:46	2.7	
18:46	3.4	
19:46	2.0	
20:46	3.4	
21:46	2.4	
22:46	2.0	
23:46	6.7	

***** 01/12/2013 *****

Time	Counts pCi/l	Flags
00:46	4.4	
01:46	4.4	
02:46	6.1	
03:46	5.4	
04:46	7.1	
05:46	4.4	
06:46	3.0	
07:46	2.7	
08:46	4.0	
09:46	2.0	
10:46	3.4	
11:46	3.7	
12:46	2.4	
13:46	3.0	
14:46	4.0	
15:46	2.4	
16:46	5.4	
17:46	3.0	
18:46	3.4	
19:46	5.4	
20:46	5.4	
21:46	5.0	
22:46	5.4	
23:46	3.4	

***** 01/13/2013 *****

Time	Counts pCi/l	Flags
00:46	2.4	
01:46	3.7	
02:46	5.0	
03:46	4.0	
04:46	4.4	
05:46	4.0	
06:46	3.0	

***** 01/13/2013 *****

Time	Counts pCi/l	Flags
07:46	2.4	
08:46	3.7	
09:46	4.0	
10:46	3.4	
11:46	5.0	
12:46	3.7	
13:46	2.7	
14:46	3.7	
15:46	5.4	
16:46	5.0	

Error Flags:

M Motion:

Inspector Signature _____

PC Software Version: 2.1.1.0

Embedded Software Version: 108