

### Continuous Radon Monitor

Model Number: 1028  
Calibration Date: 03/16/2012  
Monitor Time: 12/23/2012 21:21

Serial Number: 49745118  
CF: 2.53

### Inspector Information

Stephen R. Burns, Inc.  
P.O. Box 943  
Iowa City  
IA-52244  
Phone Number: 319-400-4844  
License Number: S00227

### Customer Information

Mann Elementary  
Room 5  
510 N.Dodge ST  
Iowa City  
IA-52245


### Site & Condition

Wind:	NA	Atmospheric Condition:	NA
Year Built:	NA	Structure Type:	NA
SqFt:	NA	Monitor Location:	NA

### Test Summary

Start Time:	12/21/2012 14:30	Overall Avg:	0.7 pCi/l
End Time:	12/23/2012 14:30	EPA Avg:	0.8 pCi/l
Measurement Interval(hr):	1.0		
Elapse Time:	2 Days 0 hrs		

### Continuous Radon Monitor



b ◆ Radon pCi/l    b ● Motion Errors    e — Temp F    e — Rel % Humidity    e — Press inHg

\*\*\*\*\* 12/21/2012 \*\*\*\*\*

Time	Counts pCi/l	Flags
15:30	0.4	
16:30	0.4	
17:30	0.0	
18:30	0.4	
19:30	0.4	
20:30	0.4	
21:30	1.6	
22:30	0.4	
23:30	0.8	

\*\*\*\*\* 12/22/2012 \*\*\*\*\*

Time	Counts pCi/l	Flags
00:30	0.8	
01:30	0.4	
02:30	0.0	
03:30	1.2	
04:30	0.0	
05:30	0.0	
06:30	0.8	
07:30	0.8	
08:30	2.8	
09:30	1.2	
10:30	2.0	
11:30	1.6	
12:30	1.6	
13:30	0.4	
14:30	0.0	
15:30	0.0	
16:30	0.0	
17:30	0.4	
18:30	0.4	
19:30	0.4	
20:30	0.0	
21:30	0.4	
22:30	0.8	
23:30	1.2	

\*\*\*\*\* 12/23/2012 \*\*\*\*\*

Time	Counts pCi/l	Flags
00:30	0.8	
01:30	1.2	
02:30	0.4	
03:30	0.4	
04:30	0.0	

\*\*\*\*\* 12/23/2012 \*\*\*\*\*

Time	Counts pCi/l	Flags
05:30	1.6	
06:30	1.2	
07:30	0.8	
08:30	0.8	
09:30	0.4	
10:30	0.0	
11:30	2.0	
12:30	0.4	
13:30	1.2	
14:30	1.6	

Error Flags:

M Motion:

Inspector Signature \_\_\_\_\_

PC Software Version: 2.1.1.0

Embedded Software Version: 108