

### Continuous Radon Monitor

Model Number: 1028  
Calibration Date: 01/27/2012  
Monitor Time: 12/17/2012 22:22

Serial Number: 49745132  
CF: 2.81

### Inspector Information

Stephen R. Burns, Inc.  
P.O. Box 943  
Iowa City  
IA-52244  
Phone Number: 319-400-4844  
License Number: S00227

### Customer Information

Iowa City School System  
Kirkwood Elementary Room 11  
1401 9th ST  
Coralville  
IA-52241

### Site & Condition

Wind:	NA	Atmospheric Condition:	NA
Year Built:	NA	Structure Type:	NA
SqFt:	NA	Monitor Location:	NA

### Test Summary

Start Time:	12/15/2012 05:30	Overall Avg:	1.0 pCi/l
End Time:	12/17/2012 05:30	EPA Avg:	1.1 pCi/l
Measurement Interval(hr):	1.0		
Elapse Time:	2 Days 0 hrs		

### Continuous Radon Monitor



b ◆ Radon pCi/l    b ● Motion Errors    e — Temp F    e — Rel % Humidity    e — Press inHg

\*\*\*\*\* 12/15/2012 \*\*\*\*\*

Time	Counts pCi/l	Flags
06:30	0.7	
07:30	0.4	
08:30	1.4	
09:30	2.1	
10:30	1.4	
11:30	1.4	
12:30	0.4	
13:30	0.4	
14:30	0.7	
15:30	0.4	
16:30	0.0	
17:30	1.1	
18:30	0.0	
19:30	1.1	
20:30	0.4	
21:30	0.7	
22:30	0.0	
23:30	0.4	

\*\*\*\*\* 12/16/2012 \*\*\*\*\*

Time	Counts pCi/l	Flags
00:30	0.7	
01:30	1.1	
02:30	0.0	
03:30	1.1	
04:30	0.7	
05:30	0.4	
06:30	2.1	
07:30	1.1	
08:30	0.0	
09:30	2.5	
10:30	1.8	
11:30	2.1	
12:30	1.4	
13:30	0.4	
14:30	1.8	
15:30	1.1	
16:30	1.8	
17:30	1.1	
18:30	1.4	
19:30	1.1	
20:30	1.1	
21:30	1.8	
22:30	1.8	
23:30	2.5	

\*\*\*\*\* 12/17/2012 \*\*\*\*\*

Time	Counts pCi/l	Flags
00:30	1.4	
01:30	0.7	
02:30	0.7	
03:30	1.4	
04:30	0.4	
05:30	1.1	

Error Flags:

M Motion:

Inspector Signature \_\_\_\_\_

PC Software Version: 2.1.1.0

Embedded Software Version: 108