

Continuous Radon Monitor

Model Number: 1028
Calibration Date: 03/16/2012
Monitor Time: 12/17/2012 22:22

Serial Number: 49745118
CF: 2.53

Inspector Information

Stephen R. Burns, Inc.
P.O. Box 943
Iowa City
IA-52244
Phone Number: 319-400-4844
License Number: S00227

Customer Information


Iowa City School System
Kirkwood Elementary Room 10
1401 9th ST
Coralville
IA-52241

Site & Condition

Wind:	NA	Atmospheric Condition:	NA
Year Built:	NA	Structure Type:	NA
SqFt:	NA	Monitor Location:	NA

Test Summary

Start Time:	12/15/2012 05:21	Overall Avg:	1.0 pCi/l
End Time:	12/17/2012 05:21	EPA Avg:	1.1 pCi/l
Measurement Interval(hr):	1.0		
Elapse Time:	2 Days 0 hrs		



***** 12/15/2012 *****

Time	Counts pCi/l	Flags
06:21	0.0	
07:21	0.4	
08:21	1.2	
09:21	0.4	
10:21	1.2	
11:21	1.6	
12:21	1.6	
13:21	0.8	
14:21	0.4	
15:21	0.4	
16:21	2.0	
17:21	0.4	
18:21	0.8	
19:21	0.0	
20:21	2.0	
21:21	2.0	
22:21	0.4	
23:21	1.2	

***** 12/16/2012 *****

Time	Counts pCi/l	Flags
00:21	1.2	
01:21	0.0	
02:21	0.4	
03:21	2.8	
04:21	0.8	
05:21	1.6	
06:21	2.0	
07:21	0.8	
08:21	0.4	
09:21	0.0	
10:21	0.8	
11:21	1.6	
12:21	0.8	
13:21	1.6	
14:21	0.8	
15:21	1.6	
16:21	0.4	
17:21	1.2	
18:21	1.2	
19:21	0.8	
20:21	2.0	
21:21	1.2	
22:21	1.2	
23:21	1.2	

***** 12/17/2012 *****

Time	Counts pCi/l	Flags
00:21	1.2	
01:21	1.6	
02:21	2.0	
03:21	0.0	
04:21	0.8	
05:21	0.0	

Error Flags:

M Motion:

Inspector Signature _____

PC Software Version: 2.1.1.0

Embedded Software Version: 108