

Continuous Radon Monitor

Model Number: 1028
Calibration Date: 09/28/2012
Monitor Time: 12/17/2012 22:22

Serial Number: 57077193
CF: 2.97

Inspector Information

Stephen R. Burns, Inc
P.O. Box 943
Iowa City
IA-52244
Phone Number: 319-400-4844
License Number: S00227

Customer Information


Iow City School System
Garner Elementary Room 108
80 Birch ST
North Liberty
IA-52317

Site & Condition

Wind:	NA	Atmospheric Condition:	NA
Year Built:	NA	Structure Type:	NA
SqFt:	NA	Monitor Location:	NA

Test Summary

Start Time:	12/15/2012 04:51	Overall Avg:	0.8 pCi/l
End Time:	12/17/2012 04:51	EPA Avg:	0.8 pCi/l
Measurement Interval(hr):	1.0		
Elapse Time:	2 Days 0 hrs		



***** 12/15/2012 *****

Time	Counts pCi/l	Flags
05:51	0.0	
06:51	0.0	
07:51	0.7	
08:51	1.0	
09:51	0.7	
10:51	0.3	
11:51	1.0	
12:51	0.3	
13:51	1.7	
14:51	0.7	
15:51	0.3	
16:51	0.7	
17:51	0.7	
18:51	1.0	
19:51	0.3	
20:51	0.3	
21:51	0.0	
22:51	0.0	
23:51	0.3	

***** 12/16/2012 *****

Time	Counts pCi/l	Flags
00:51	0.3	
01:51	1.0	
02:51	0.7	
03:51	0.7	
04:51	1.0	
05:51	3.4	
06:51	2.0	
07:51	2.0	
08:51	1.7	
09:51	1.0	
10:51	1.3	
11:51	1.7	
12:51	0.7	
13:51	0.7	
14:51	0.0	
15:51	0.0	
16:51	0.0	
17:51	0.0	
18:51	0.7	
19:51	0.3	
20:51	0.3	
21:51	1.0	
22:51	0.0	

***** 12/16/2012 *****

Time	Counts pCi/l	Flags
23:51	2.7	

***** 12/17/2012 *****

Time	Counts pCi/l	Flags
00:51	1.7	
01:51	0.3	
02:51	0.3	
03:51	0.3	
04:51	0.7	

Error Flags:

M Motion:

Inspector Signature _____

PC Software Version: 2.1.1.0

Embedded Software Version: 108