

Continuous Radon Monitor

Model Number: 1028
Calibration Date: 09/28/2012
Monitor Time: 4/15/2013 23:23

Serial Number: 59237107
CF: 2.88

Inspector Information

Stephen R. Burns Inc.
P.O. BOX 943
Iowa City
IA-52244
Phone Number: 319-400-4844
License Number: S00227

Customer Information


ICCSA
Coralville Room 9
501 6th ST
Coralville
IA-52241

Site & Condition

Wind: NA
Year Built: NA
SqFt: NA
Atmospheric Condition: NA
Structure Type: NA
Monitor Location: NA

Test Summary

Start Time: 04/13/2013 04:36
End Time: 04/15/2013 04:36
Measurement Interval(hr): 1.0
Elapse Time: 2 Days 0 hrs
Overall Avg: 4.2 pCi/l
EPA Avg: 4.4 pCi/l



***** 04/13/2013 *****

Time	Counts pCi/l	Flags
05:36	2.1	
06:36	2.4	
07:36	2.8	
08:36	4.2	
09:36	4.2	
10:36	2.1	
11:36	1.4	
12:36	2.1	
13:36	2.4	
14:36	3.5	
15:36	5.2	
16:36	4.9	
17:36	3.5	
18:36	4.2	
19:36	2.8	
20:36	4.5	
21:36	4.2	
22:36	3.5	
23:36	5.2	

***** 04/14/2013 *****

Time	Counts pCi/l	Flags
00:36	2.4	
01:36	3.8	
02:36	3.8	
03:36	4.5	
04:36	2.8	
05:36	4.2	
06:36	8.0	
07:36	5.6	
08:36	5.9	
09:36	3.5	
10:36	6.6	
11:36	5.2	
12:36	4.5	
13:36	8.3	
14:36	3.8	
15:36	6.6	
16:36	6.6	
17:36	2.8	
18:36	4.5	
19:36	4.2	
20:36	6.6	
21:36	3.8	
22:36	4.5	

***** 04/14/2013 *****

Time	Counts pCi/l	Flags
23:36	4.5	

***** 04/15/2013 *****

Time	Counts pCi/l	Flags
00:36	4.5	
01:36	3.8	
02:36	4.5	
03:36	3.5	
04:36	3.1	

Error Flags:

M Motion:

Inspector Signature _____

PC Software Version: 2.1.1.0

Embedded Software Version: 108