

Continuous Radon Monitor

Model Number: 1028
Calibration Date: 09/28/2012
Monitor Time: 12/17/2012 15:15

Serial Number: 49745178
CF: 2.84

Inspector Information

Stephen R. Burns, Inc.
P.O. Box 943
Iowa City
IA-52244
Phone Number: 319-400-4844
License Number: S00227

Customer Information


Iowa City School District
Coralville Central Elementary Rm 207
501 6th ST
Coralville
IA-52241

Site & Condition

Wind: NA
Year Built: NA
SqFt: NA
Atmospheric Condition: NA
Structure Type: NA
Monitor Location: NA

Test Summary

Start Time: 12/15/2012 06:03
End Time: 12/17/2012 06:03
Measurement Interval(hr): 1.0
Elapse Time: 2 Days 0 hrs
Overall Avg: 6.0 pCi/l
EPA Avg: 6.3 pCi/l



***** 12/15/2012 *****

Time	Counts pCi/l	Flags
07:03	2.1	
08:03	5.3	
09:03	2.8	
10:03	6.3	
11:03	5.3	
12:03	5.3	
13:03	5.3	
14:03	6.7	
15:03	8.1	
16:03	6.0	
17:03	9.2	
18:03	9.2	
19:03	6.3	
20:03	6.0	
21:03	5.6	
22:03	6.3	
23:03	4.6	

***** 12/16/2012 *****

Time	Counts pCi/l	Flags
00:03	3.9	
01:03	7.0	
02:03	6.0	
03:03	7.0	
04:03	7.4	
05:03	5.6	
06:03	4.6	
07:03	3.9	
08:03	4.6	
09:03	7.7	
10:03	4.6	
11:03	4.6	
12:03	8.1	
13:03	6.7	
14:03	5.3	
15:03	7.0	
16:03	6.3	
17:03	6.7	
18:03	5.3	
19:03	6.0	
20:03	4.9	
21:03	4.6	
22:03	4.6	
23:03	4.9	

***** 12/17/2012 *****

Time	Counts pCi/l	Flags
00:03	7.4	

***** 12/17/2012 *****

Time	Counts pCi/l	Flags
01:03	8.4	
02:03	7.4	
03:03	8.4	
04:03	4.6	
05:03	6.7	
06:03	7.7	

Error Flags:

M Motion:

Inspector Signature _____

PC Software Version: 2.1.1.0

Embedded Software Version: 108