

Continuous Radon Monitor

Model Number: 1028
Calibration Date: 09/28/2012
Monitor Time: 2/18/2013 20:20

Serial Number: 49745178
CF: 2.84

Inspector Information

Stephen R. Burns, Inc.
P.O. Box 943
Iowa City
IA-52244
Phone Number: 319-400-4844
License Number: S00227

Customer Information

Borlaug Elementary
Room 1105
1000 Kennedy PKWY
Coralville
IA-52241


Site & Condition

Wind: NA
Year Built: NA
SqFt: NA
Atmospheric Condition: NA
Structure Type: NA
Monitor Location: NA

Test Summary

Start Time: 02/16/2013 04:25
End Time: 02/18/2013 04:25
Measurement Interval(hr): 1.0
Elapse Time: 2 Days 0 hrs
Overall Avg: 2.0 pCi/l
EPA Avg: 2.1 pCi/l

Continuous Radon Monitor



b ◆ Radon pCi/l b ● Motion Errors e — Temp F e — Rel % Humidity e — Press inHg

***** 02/16/2013 *****

Time	Counts pCi/l	Flags
05:25	1.1	
06:25	0.0	
07:25	0.4	
08:25	2.1	
09:25	1.4	
10:25	1.4	
11:25	0.7	
12:25	1.4	
13:25	1.1	
14:25	2.8	
15:25	0.7	
16:25	1.1	
17:25	0.4	
18:25	2.5	
19:25	1.4	
20:25	0.7	
21:25	2.5	
22:25	1.4	
23:25	0.7	

***** 02/17/2013 *****

Time	Counts pCi/l	Flags
00:25	3.2	
01:25	2.8	
02:25	1.8	
03:25	1.1	
04:25	3.5	
05:25	2.1	
06:25	3.5	
07:25	3.2	
08:25	2.1	
09:25	4.2	
10:25	3.2	
11:25	3.9	
12:25	1.8	
13:25	2.5	
14:25	1.8	
15:25	2.8	
16:25	1.8	
17:25	1.8	
18:25	1.1	
19:25	4.6	
20:25	3.5	
21:25	1.1	
22:25	0.7	

***** 02/17/2013 *****

Time	Counts pCi/l	Flags
23:25	1.4	

***** 02/18/2013 *****

Time	Counts pCi/l	Flags
00:25	0.7	
01:25	1.4	
02:25	3.2	
03:25	2.5	
04:25	3.5	

Error Flags:

M Motion:

Inspector Signature _____

PC Software Version: 2.1.1.0

Embedded Software Version: 108